

**Accurate, Reliable
and Affordable
Single Parameter Analysis**



ChemScan mini Analyzers

The single parameter in-line analyzer family uses years of ChemScan experience and proven technology to provide reliable and accurate analysis of water and waste water. This device has been designed from the ground up to reduce maintenance requirements. It includes large ID sample tubing to minimize plugging and needs reagent refills only quarterly.

CAPABILITIES

- Continuous, real-time analysis of constant flow sample stream
- Isolated analog output

FEATURES

- Long life LED light source
- Low maintenance
- Large I.D. flow paths
- Simple field adjustable calibration
- Direct diode detection
- Sealed electronics enclosure
- Auto cleaning and zeroing
- No lamp replacement or alignment required
- No filtration required
 - When TSS < 150 mg/L
 - After secondary clarifier
- Filter accessories available


BENEFITS

- High reliability
- Low capital cost
- High accuracy

www.ChemScan.com

Email: info@chemscan.com
800-665-7133 (toll-free in U.S.A. and Canada)
1-262-717-9500 (U.S.A. and international)

ACCESSORIES



Sample Extraction Accessory

Provides a pressurized sample to the ChemScan mini analyzer, where NTU is less than 60 and TSS is less than 150 mg/L

TSS - Total Suspended Solids
NTU - Nephelometric Turbidity Units



ChemScan Cartridge Filter Wand

For high-solids applications. No pressurized air, water or chemicals required for cleaning.



ChemScan mini Outdoor Enclosure

A turnkey solution for mounting the ChemScan analyzer and related items.



Submersible Pump

1.3" Maximum diameter solids
Weight: 20-30 lbs.
Power: 1/4 - 3/4 HP, 120 VAC 60 Hz
Power Cable: 20 feet



Deck-Mounted Self-Priming Pump

1/3 - 1/2 HP
Weight: 40 lbs
Mounting: Base

General (Common to all minis)

Accuracy:	2% of value or 2x detection limit (whichever is greater)
Environment:	5 - 50 degrees C
Power:	100 - 240 VAC, 50 W
Enclosure:	NEMA 4X
Safety Approval:	CSA-US
Relay Contacts:	1 SPDT Concentration, 1 SPDT Programmable
Serial Interface:	Serial, RS-232, Modbus RTU
Analog Output:	Isolated 4-20 mA
Sample:	0.5 - 1 Liter/analysis, pressure to 10 psi (UV-254 Continuous)

ChemScan mini oP

Range (as PO ₄):	0.1 - 9.0 mg/L (Method 1005), 0.3 - 18.0 mg/L (Method 1006)
Range (as PO ₄ -P):	0.03 - 3.0 mg/L (Method 1003), 0.1 - 6.0 mg/L (Method 1004)
Cycle Interval:	5 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly

ChemScan mini oP XR

Range (as P):	0.1 - 20.0 mg/L (Method 1069)
Range (as PO ₄):	0.3 - 60.0 mg/L (Method 1070)
Cycle Interval:	5 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

ChemScan mini LoP

Range (as PO ₄):	0.02 - 3.0 mg/L (Method 1071)
Range (as PO ₄ -P):	0.003 - 1.00 mg/L (Method 1034)
Cycle Interval:	8 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

ChemScan mini LowAm

Range (as N):	0.01 - 10.0 mg/L (Method 1066)
Cycle Interval:	15 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

CHLORAMINATION SUITE

ChemScan mini FreeAm

Range (as N):	0.01 - 2.00 mg/L (Method 1036)
Cycle Interval:	18 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every month, pump kit yearly*

ChemScan mini Mono

Range (as N):	0.05 - 10.0 mg/L (Method 1035)
Cycle Interval:	10 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

DRINKING WATER SUITE

ChemScan mini Mn

Range:	0.02 - 8.0 mg/L (Method 1063, 1064)
Cycle Interval:	15 min. (1064) 10 min. (1063) to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

ChemScan mini Fe

Range:	0.01 - 5.0 mg/L (Method 1039)
	0.02 - 20.0 mg/L (Method 1037)
Cycle Interval:	8 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

WASTEWATER DISINFECTION SUITE

ChemScan mini Sulfite

Range:	0.01 - 4.0 mg/L (Method 1068)
Cycle Interval:	5 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every month, pump kit yearly*

ChemScan mini LowChlor

Range (as CL ₂):	0.005 - 2.00 mg/L (Method 1030)
Cycle Interval:	5 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every month, pump kit yearly*

ChemScan mini CrVI

Range:	0.03 - 5.0 mg/L (Method 1040)
Cycle Interval:	12 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

ChemScan mini LowCrVI

Range:	1 - 1000 µg/L (Method 1041)
Cycle Interval:	12 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

ChemScan mini Silica

Range:	0.05 - 15.0 mg/L (Method 1058)
Cycle Interval:	7 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 6 months, pump kit yearly*

ChemScan mini Ni

Range:	0.05 - 6.0 mg/L (Method 1057)
Cycle Interval:	8 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

ChemScan mini UV254

Range (as N):	0.1 - 100%T
Cycle Interval:	Continuous
Sample:	2 - 10 psi continuous flow
Maintenance:	Replace zero/clean solution

ChemScan mini Peracetic Acid (PAA)

Range:	0.015 - 5.0 mg/L (Method 1073)
Cycle Interval:	5 minutes to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 4 weeks, pump kit yearly*

ChemScan mini Cu

Range:	0.02 - 6.0 mg/L (Method 1065)
	0.001 - 2.00 mg/L (Method 1056)
	0.05 - 6.0 mg/L (Method 1027)
Cycle Interval:	4 min. (1027) 5 min. (1065) 4 min. (1056) to 9999 minutes (field programmable)
Maintenance:	Reagent replacement every 3 months, pump kit yearly*

CHLORAMINATION ANALYZER

ChemScan mini ChlorAm

Range:	Free Ammonia 0.025 - 2.00 mg/L
	Total Ammonia 0.02 - 3.00 mg/L
	Monochloramine 0.02 - 5.00 mg/L
	Ratio - Calculated using Total Ammonia and Monochloramine
Cycle Interval:	18 minutes to 9999 minutes with 9 minute updates
Maintenance:	Reagent replacement every month, pump kit yearly*

* Based on default cycle time